

Supplementary Material

1 Supplementary Table

Supplementary Table 2. CRP test results obtained in relation with the rapid CRP Eurolysed results.

| CRP Eurolyser results grouped | CRP Rapid test results | | | |
|-------------------------------|------------------------|--------------------|--------------------|------------------|
| | <10 mg/l [n (%)] | 10-40 mg/l [n (%)] | 40-80 mg/l [n (%)] | >80 mg/l [n (%)] |
| <10 mg/l (n=142) | 120 (84.5) | 22 (15.5) | 0 (0) | 0 (0) |
| 10-40 mg/l (n=56) | 3 (5.4) | 48 (85.7) | 5 (8.9) | 0 (0) |
| 40-80 mg/l (n=31) | 0 (0) | 7 (22.6) | 17 (54.8) | 7 (22.6) |
| >80 mg/l (n=47) | 0 (0) | 1 (2.1) | 4 (8.5) | 42 (89.4) |
| Total (n=276*) | 123 (44.6) | 78 (28.3) | 26 (9.4) | 49 (17.8) |

* CRP Eurolyser result was missing for one sample.

kappa (SE) = 0.736 (0.033)